

**HIGH TEMPERATURE
STANDARD PLATINUM
RESISTANCE THERMOMETER
MODEL 96178**



Isotech North America
158 Brentwood Drive, Unit 4
Colchester, VT 05446

Phone: (802)-863-8050
Fax: (802)-863-8125

www.isotechna.com
sales@isotechna.com

The company is always willing to give technical advice and assistance where appropriate. Equally, because of the programme of continual development and improvement, we reserve the right to amend or alter characteristics and design without prior notice. This publication is for information only

CONTENTS

	PAGE NO
1. Guarantee	3
2. Introduction	4
4. Returning your 96178 to ISOTECH	5
5. Stability Evaluation	6 -12
6. Glossary of Metrological Terms	13 -14
 Packing Instruction for Return of Thermometers 909-20-01	 15

GUARANTEE

This instrument has been manufactured to exacting standards and is guaranteed for twelve months against electrical break-down or mechanical failure caused through defective material or workmanship, provided the failure is not the result of misuse. In the event of failure covered by this guarantee, the instrument must be returned, carriage paid, to the supplier for examination and will be replaced or repaired at our option.

FRAGILE CERAMIC AND/OR GLASS PARTS ARE NOT COVERED BY THIS GUARANTEE

INTERFERENCE WITH, OR FAILURE TO PROPERLY MAINTAIN THIS INSTRUMENT MAY
INVALIDATE THIS GUARANTEE

RECOMMENDATION

The life of your **ISOTECH** Instrument will be prolonged if regular maintenance and cleaning to remove general dust and debris is carried out.

We recommend this instrument to be re-calibrated annually.

Serial No:.....

Date:.....



Isotech North America
158 Brentwood Drive, Unit 4
Colchester, VT 05446

Phone: (802)-863-8050
Fax: (802)-863-8125

www.isotechna.com
sales@isotechna.com

OPERATING INSTRUCTIONS

INTRODUCTION

You or your company has just paid a considerable sum of money to purchase this 96178.

Although the materials to make the 96178 are very costly, it has also taken up to 1000 hours to prepare the 96178 for your use.

It is not a product that can be assembled, tested and sold, it is produced, calibrated, aged and recalibrated, until its characteristics are stable enough to meet the exacting needs of ITS 90.

The weeks of work that go into its production make each 96178 more than another product. Each 96178 already has a character and a history before it leaves us. Please look after it. Regard yourself as its custodian rather than its owner.

1. Always keep the 96178 in its case when not in use.
2. When in use, support the handle.
3. Cool and store the 96178 in the same place as you normally use it.
4. Each time before you use the 96178 clean off all traces of grease by using a chlorinated solvent.
5. Even go so far as using gloves to handle the 96178, and keep the gloves clean.
6. Quartz is glass. It is a supercooled liquid. At 800°C and above, your 96178 will bend and bow if you do not support it along its complete length. At high temperatures I recommend that the 96178 be housed inside a close fitting recrystallised Alumina closed ended tube, which has been pre-fired to 1000°C.
7. Quartz is transparent in two senses of the word. At temperatures above 700°C metallic vapours can pass through the quartz and attack the pure platinum sensing element.

Isotech have developed a product which can be attached to the 96178 to prevent this happening a 960 Ioniser. A Model 960 Ioniser is provided free with each 96178.

RETURNING YOUR 96178 TO ISOTECH

Due to the fragility of 96178 we strongly recommend that it is not transported unless personally hand carried.

In the event that the unit must be shipped by courier then the following instructions must be adhered to

REPACKING INSTRUCTIONS

Place a small foam spacer under the end of the sheath and one under the handle.

Then place two larger pieces over the end of the sheath.

Preparing the boxes

Place the two foam blocks over the closed case and lower into inner box. Pad out ends with polystyrene chips to stop the unit from moving - seal the box, lower it into the outer box and fill with chips, then lift the inner box slightly so the polystyrene chips completely surround it. Place the protective sheath on top of the chips and seal the box.

Always remember to label the box thoroughly with fragile and this way up labels and ensure you have adequate insurance cover.

Your unit should now be ready to send safely.

STABILITY EVALUATION

INTRODUCTION

The 96178 is one of a new generation of thermometers designed at Isotech. It is a high temperature thermometer working up to 1000°C and being ¼ ohms at 0 °C.

It's main radical and new features are:

1. The 96178 can be supplied pre-sealed with an inert gas containing 10% oxygen or with a valve built into the handle so the user can allow the thermometer to breath.
2. The quartz used is a new and more pure material.
3. The internal construction is made of many small parts to eliminate stem conduction.
4. The internal structure contains a platinum light scattering disc to reflect radiant energy.

In this report the first production unit was cycled to the silver point to test the reproducibility of the W value at silver, and to check the stability of R_{TPW} .

METHOD

The 96178 thermometer, like all high temperature thermometers will change its characteristics if subjected to step changes of temperature. The way in which the thermometer is temperature cycled is therefore very important.

In this series of tests the following strategy was adopted:-

The thermometer was placed in a pre-warming/annealing furnace at 400°C. The furnace temperature was then slowly raised to 850°C (The apparatus used was the Isotech Dual Furnace which has been pre-programmed with ramp-rates suitable for pre-warming the High Temperature Thermometers). From this furnace, the thermometer was transferred to a heat pipe furnace in which was placed a silver cell. The silver had previously been melted and partially frozen.

After 30 minutes in the cell, during which time and after stabilisation 4 to 6 readings were recorded, the thermometer was transferred back to the annealing furnace, still at 850°C.

The annealing furnace was then cooled to 425°C overnight and R_{TPW} measured the next morning.

Other thermometers of proven good stability were cycled with the 96178 to add further confidence that the method gave stable results.

RESULTS

The following results were obtained during two temperature cycles to the silver point. Mean values are tabulated.

Serial No. 96178/1 is the new thermometer.

Serial No. 962/148 the best thermometer from my store of standards - it also "breathes" i.e. the air inside the stem is not sealed in.

Ohms or ratio	962/148	96178/1
Initial R_{TPW}	0.2362608	0.2100675
R Ag	1.0127804	0.9004547
R TPW	0.2362600	0.2100668
W Ag	4.2867197	4.2865160
W Ga	1.1181537	1.1181429
R Ag ₍₂₎	1.0127804	0.9004532
R TPW ₍₂₎	0.2362602	0.2100667
W Ag ₍₂₎	4.2867160	4.2865109
W Ga ₍₂₎	1.1181541	1.1181439

DISCUSSION OF RESULTS

High temperature thermometers take a long while - many hundred, if not thousands of hours - to stabilise completely. Thermometer 962/148 is some 3 to 4 years old and has proved to be our most stable thermometer.

96178/1 was stabilised only overnight after its arrival from manufacture; even so the stability of the new thermometer is remarkable.

There is always an initial shift in the characteristics of a high temperature thermometer during its first cycle to the silver point. In this instance, both thermometers became more fully annealed, and so the R_{TPW} was reduced by 0.7 & 0.8 mK; then, between the first and second calibrations the R_{TPW} remained stable at 0.1 and 0.2 mK.

Some high temperature thermometers exhibit better stability in the reproducibility of the resistance at the silver point, whilst others keep better W silver stability. In this evaluation we can look at both R silver and W silver.

The R silver reproducibility of Serial No. 962/148 was perfect -a very unusual - if not unique - situation.

96178/1 thermometer reproduced R silver by 1.5 mK, still exceptionally good.

962/148 reproduced W silver to within the equivalent of about 1 mK

96178/1 to within the equivalent of about 1.5 mK.

These are both exceptionally good results.

To put the above results into context; N.P.L. would permit a shift of up to 5 mK in R_{TPW} and a spread of 20 mK in W silver during a full calibration cycle.

CONCLUSION

The 96178/1 thermometer has shown itself to be as stable as our very best standard thermometer after only an overnight anneal.

FURTHER WORK

As more thermometers are produced and tested further reports will be issued to confirm the results above.

Uncertainties vary depending on the temperature range of calibration.

Isotech's 909 and 962 thermometers are sufficiently stable to conform to uncertainties translated overleaf. Models 670 and 96178 can be calibrated as tabulated overleaf or at an extra cost to the uncertainties of Issue 24 of our schedule tabulated on page 12.

The enclosed uncertainties apply on condition that the measurements of resistance at the triple point of water are reproducible within the uncertainty given at that temperature. Uncertainties for thermometers which do not meet this requirement will be increased by a factor of 3 or 10, as appropriate. For the time being, metal-sheathed thermometers will be calibrated by comparison with standard thermometers at the boiling points of water and sulphur. Calibrations at the triple point of argon are obtained from comparisons with standard thermometers in a bath of liquid nitrogen.

United Kingdom Accreditation Service

CALIBRATION LABORATORY
 No. 0175

SCHEDULE



National Accreditation of Measurement
 Applications
 Sampling

Address of permanent laboratory Isothermal Technology Ltd Pine Grove Southport Merseyside PR9 6AG Telephone : Southport (01704) 544930/544931 Fax : 01704 544799	Category 0 Permanent Laboratory Calibration performed on permanent laboratory premises APPROVED SIGNATORIES Head of Laboratory: Mr J P Towner Deputy: Mr D J Ayres Mr A S Dundell, Mr D Southworth, Mr N Davies, Mr A Orme Issue No: 13 Date: 24 February 1987
--	---


Measured Quantities for which UKAS has granted NAMAS Accreditation

Item	Measured Quantity Instrument or Gauge	Range	Best Measurement Capability Expressed as an Expanded Uncertainty (k=2)			
	TEMPERATURE					
1	Platinum Resistance	0 to 1100 °C Above 100 to 1200 °C	1 K 2 K			
2	Other thermocouples	-185 °C 20 to 250 °C Above 250 to 600 °C Above 600 to 900 °C Above 900 to 1100 °C Above 1100 to 1200 °C	0.5 K 0.5 K 1 K 2 K 3 K 4 K			
3	Thermocouples with extension cables	-25 to 200 °C	1 K			
4	Thermocouples without extension cables					
Uncertainty (k)						
	Temperature (°C)	Range 1	Range 2	Range 3	Range 4	Range 5
	TP (Triple Point)		1.0 uK	1.0 uK	1.0 uK	1.0 uK
	TP (Melting Point)	2.0 K	2.0 K	2.0 K	2.0 K	2.0 K
	TP (Boiling Point)	1.0 K	1.0 K	2.0 K	2.0 K	11.0 K
	BP (Boiling Point)	2.0 K				
	BP (Melting Point)		2.0 K			
	BP (Triple Point)		1.5 mK	1.5 uK	5.0 K	10.0 K
	BP (Boiling Point)			1.5 mK	5.0 K	10.0 K
	BP (Melting Point)				10.0 K	55.0 K
	BP (Boiling Point)					40.0 K

Note: TP = Triple Point, BP = Boiling Point,
 MP = Melting Point, BP = Boiling Point

*The Expanded Uncertainty is given for k=2, providing a level of confidence of approximately 95% issued by the United Kingdom Accreditation Service

Sheet 1 of 3

 0110 Calibration performed on accordance with ISO 9001:2008	Schedule of Accreditation (DRAFT) Issued by United Kingdom Accreditation Service 21 - 47 Edge Avenue, Feltham, Middlesex, TW9 3JUN, U.K.
	Isothermal Technology Ltd Issue No: 026 Issue date: January 2002

DETAIL OF ACCREDITATION

Measured Quantity Instrument or Gauge	Range	Best Measurement Capability Expressed as an Expanded Uncertainty (k=2)	Remarks		
TEMPERATURE					
Water in thermocouple Calibrated by comparison	200C to 600C above 600C to 1100C above 1100C to 1300C	0.20C 0.30C 0.30C			
Oil in thermocouple	220C to 600C	0.40C			
Gold/Platinum thermocouple Calibrated at fixed point	910 to 1300C 400 to 600C	0.10C 0.050C			
Oil in thermocouple	-140C above 0C to 300C 0C to 300C above 300C up to 600C above 600C up to 800C above 800C to 1100C above 1100C to 1300C	0.20C 0.250C 0.10C 0.200C 0.300C 0.40C 0.50C 0.50C			
Conductivity cell calibration	100 to 200C	1%			
Analysis of trace elements Zinc Food and feed stuffs					
Uncertainty (U)					
Temperature	Class 1	Class 2	Class 3	Class 4	Class 5
DP Storage	0.5	1mk	1mk	2mk	3mk
TP Memory	0.5 to 1.5	0.5mk	1mk	1mk	2mk
TP Warm	0.5	0.5mk	1mk	1mk	2mk
MP of 2.0s	20.00s	1.5mk			4mk
TP Input	150.000	1mk		1mk	2mk
TP In	251.000	1mk		1mk	2mk
TP Out	415.227			12mk	25mk
TP Alarm 1	560.000				4mk
TP Alarm 2	601.000				11mk
Notes: TP = Trip Point MP = Melting Point	MP = Melting Point SP = Spiking Point				
Calibration by comparison	below -60C down to -80C -60C up to -40C above 20C up to 100C above 100C up to 400C above 200C up to 600C above 400C up to 800C	1.0mk 4.0mk 0.3mk 0.5mk 0.5mk 0.5mk			
Fixed points	Melting point of 400C Water triple point (0.01C)	1.0mk 0.12mk			
					The digital output (DO) is optional for the Kelvin (C) temperature sensor.

GLOSSARY OF METROLOGICAL TERMS

ACCURACY OF MEASUREMENT

The closeness of the agreement between the result of a measurement and the (conventional) true value of the measurand*

REPRODUCIBILITY OF MEASUREMENTS

The closeness of the agreement between the results of measurements of the same measurand, where the individual measurements are carried out changing conditions such as:

- method of measurement
- observer
- measuring instrument
- location
- conditions of use
- time

NOTES

1. A valid statement of reproducibility requires specification of the conditions changed.
2. Reproducibility may be expressed quantitatively in terms of the dispersion of the results.

UNCERTAINTY OF MEASUREMENT

An estimate characterising the range of values within which the true value of a measurand lies.

NOTE

Uncertainty of measurement comprises, in general, many components. Some of these components may be estimated on the basis of the statistical distribution of the results of series of measurements and can be characterised by experimental standard deviations. Estimates of other components can only be based on experience or other information.

STABILITY

The ability of a measuring instrument to maintain constant its metrological characteristics.

DRIFT

The slow variation with time of metrological characteristic of a measuring instrument.

*Measurand = a quantity subjected to measurement

